Critec[®] ... and life gets guaranteed!

Laboratory Deep Freeze Storage Cabinet

Internally spark resistant, *Static cooling* (also in accordance to DIN 13277)

· · · · · · · · · · · · · · · · · · ·	External Dimensions:	W = 790 mm D = 1000 mm H = 2070 mm
E	Internal Dimensions:	W = 660 mm D = 870 mm H = 1500 mm
	Capacity:	900 I
anti- bacterial	Temperature range:	-5°C to -22°C

Housing

made of **Stainless steel**. (Backplane, as well as top and bottom side of the cabinet of galvanised sheet steel, antibacterial coated). With 4 feet, height adjustable

Interior space

Stainless steel, antibacterial coated cleaning friendly by rounded corners, slippery surface in the interior, meets highest hygiene requirements

Isolation

Polyurethane 60mm, free of CFCS

Cooling

Static cooling, with foamed evaporator inside the side walls

Door

Solid door, supplied as standard with right-hand hinge, also available with left-hand hinge at no extra cost. Grease resistant magnetic seal. **Optional:** with **door lock**

Interior fittings 8 pcs. Grating-type shelves white coated (dim.: 600x400 mm), 2 shelves per layer

Optional: stainless steel drawer, without lengthwise divider (max. 13 drawers possible). Drawer interior dimensions: 571x764x77 mm On castors, telescopic guide with locking device

Technical indication:

This refrigerator is classified as **"internally spark resistant"** and have no electrical components in the interior, reducing the risk of fire or explosion. (The class of explosion protection provided by this equipment meets Federal German safety regulations for laboratory equipment DGUV Information 213-850. The appropriate regulations must be observed in other countries). **The Device complies with the ATEX directive: 2014/34/EU**

tritec ® Gesellschaft für Labortechnik und Umweltsimulation mbH Hüttenstraße 9 D-30165 Hannover

Homepage
E-mailwww.tritec-klima.com
info@tritec-klima.dePhone++49-511/3523508Fax++49-511/3521715technischeÄnderungen vorbehalten



TC 1019-ex

Seite 1 von 5



Laboratory Deep Freeze Storage Cabinet TC 1019-ex Internally spark resistant, *Static cooling* (also in accordance to DIN 13277)

Control system located above the door

Touchscreen temperature controller, TC 2015



Working range from -5°C to -22°C

Temperature accuracy +/- 3,5°C after stabilising (with 50% filling)

Language in German / English adjustable

- Touchscreen Display, size 12,5 cm (5")
- PT1000 temperature sensor for better temperature control accuracy
- Digital temperature setting and display with an accuracy of 0.1 ° C
- Power supply of 100-240V, 50-60Hz
- Password protected to prevent manipulations
- Data import/export via USB or Ethernet

Alarm

- Noticeable optical alarm signalling by red / blue flashing screen
- Audible alarm
- Mute the audible alarms, with continuous warning indication on the display and automatic return after 3 minutes when the alarm situation still exists
- Alarm acknowledgment with password protection

Acoustic and optical alarm indication when

- High and low temperature (variably adjustable)
- Door open alarm, (adjustable)
- Power failure
- Temperature sensor failure
- Defrost-time transgression (variable adjustable)
- Optional: cooling machine monitoring with overheat protection
- Alarm memory (not deleted) for the last 100 alarm conditions, with the possibility of review on the display and archiving the data via USB or Ethernet. Memory can not be deleted.

Event memory

In the event log all the operations of the controller are saved, such as:

- Doorways
- Parameter changes
- Cooling equipment function, etc

- Storage for approximately 4 weeks, with the possibility of review on the display and archiving the data via USB or Ethernet.

Graphical display of the temperature profile in the display

Illustration of the actual value and if requested of the product sensor on the display and archiving via USB or Ethernet. View period freely selectable.

tritec ® Gesellschaft für Labortechnik und Umweltsimulation mbH Hüttenstraße 9 D-30165 Hannover

Homepage
E-mailwww.tritec-klima.com
info@tritec-klima.dePhone++49-511/3523508Fax++49-511/3521715technischeÄnderungen vorbehalten





Laboratory Deep Freeze Storage Cabinet TC 1019-ex Internally spark resistant, *Static cooling* (also in accordance to DIN 13277)

Network Connection

2 Ethernet ports 10/100 Mbps available. This interface can be accessed directly on the log data of the controller to import in the software.

NO/NC contact

Alarm signal time delayed (adjustable)

- Language: English / German, other languages on request
- Maintenance message: Notification that maintenance should be carried out.
- Test facilities for Alarm limits and battery charge status
- In the housing of the temperature regulator there are all relevant components for using the cabinet. Plug-in connectors of the electrical supply pipes enable a **service friendly exchange**.

Optional

- Battery Module

In case of power failure self-contained supply of the temperature controller with a powerful battery. All control functions are saved for about 72 hours.

- 4...20 mA Output
- WLAN-Module
- Interface, RS 485 (galvanically isolated)
- <u>Software</u>

This software enables you to check at a glance several cooling units. It was designed to be user friendly and enables you to process your data in an intuitive manner.

- Alarm system
 - Independent product sensor for reference measurement
 - Product protection against low temperatures. By dropping below the setpoint value, the cooling machine will switched off
 - Battery module as a standard feature

Security barrier (EN 50014 and EN 50020) for temperature sensor (according to ATEX directive)

Refrigerating unit

fully hermetically sealed, forced ventilation, fitted on vibration - absorbing mounts (ambient temperature max. 25°C), low noise (54 dB(A)), energy saving compressor with high quality vaporisation system. Refrigerant: R290 or similar

Defrost

Manually. With Dew water collector tank, disposed below the cabinet.

Electrical Data

Power supply230 V/50 Hz /single phaseOptional:60HzPower input0,85 kWFuse16 APower cable:2,0 m with schuko plug

Packing details (palletized) Dimensions: approx. 90x115x224 cm Net weight: 205 kg

Gross weight: 230 kg Country of Origin: European Union Customs clearance code: 8418 4080

tritec ® Gesellschaft für Labortechnik und Umweltsimulation mbH Hüttenstraße 9 D-30165 Hannover

Homepagewww.tritec-klima.comE-mailinfo@tritec-klima.dePhone++49-511/3523508Fax++49-511/3521715technischeÄnderungen vorbehalten





Laboratory Deep Freeze Storage Cabinet TC 1019-ex Internally spark resistant, *Static cooling* (also in accordance to DIN 13277)


Special Equipment and Accessories:



GSM Modul

Connecting to the potential-free output. In case of an alarm either a message or a call will be sent automatically. Archiving of 1000 phone numbers is possible. The GSM module is equipped with a rechargeable battery. Automatic alert via SMS when the credit has been used on the SIM card. 6 units can be connected per module. The SIM card is not included

Round Chart recorder





Door lock including 2 keys <u>Optional</u>: electronic door lock, via temperature controller TC 2015



Cable port with cover (approx. 40 mm \emptyset).

Cable port with PG-gland (approx. 13mm \oslash)

for example, to create access for operator measurement lines, etc.

Optional: with separate sensor kind and version as desired by the customer

Homepage
E-mailwww.tritec-klima.com
info@tritec-klima.dePhone++49-511/3523508Fax++49-511/3521715technischeÄnderungen vorbehalten



Criter[®] ... and life gets guaranteed!

Laboratory Deep Freeze Storage Cabinet

Internally spark resistant, *Static cooling* (also in accordance to DIN 13277)



TC 1019-ex

Qualifications

DQ (Design Qualification)

Definition: Documented proof that the quality-related, GMP-related requirements has been adequately addressed in the design of equipment, including buildings, premises and auxiliary equipment

The user-requirement profiles (specifications) are documented and confirmed by us. On request, a specification can be created by us.

IQ (Installation Qualification)

Definition: Documented proof that critical equipment and systems have been delivered and installed in accordance with the set requirements and government regulations.

The IQ documentation is worked out by us especially for the delivered machine and is made available to you. The IQ documentation must be carried out by the customer itself.

OQ (Operational Qualification)

Definition: Documented proof that critical equipment and systems in accordance with the set requirements in the whole operating range are working as intended in accordance with predetermined limits.

The OQ documentation is worked out by us especially for the delivered machine and is made available to you. The OQ documentation must be carried out by the customer itself.

CQ (Calibration Qualification) according to DIN 13277:2022-05

Definition: Documented proof that critical measuring equipment in the intended range in accordance with predetermined tolerances operate reliably under current operating conditions

Verifying the temperature in the unloaded cooling unit (after reaching the steady state)

1 temperature on 3 measuring levels with 5 measuring points each

(Measurement with calibrated PT 1000 sensors). Test time 4 hours, then open door for 60 seconds. During this time, the limit values specified in DIN 13277:2022-05 must not be exceeded. Repeat the door opening after one hour.

The temperature measurements are carried out on our premises. The evaluation of the measurements, including graphical representation, is made in written form. The values must not exceed the limit values specified in DIN 13277:2022-05. (Other measuring methods possible on request)

PQ (Performance-Qualification) according to DIN 13277:2022-05

Definition: Documented proof that critical equipment and systems in accordance with the set requirements in the whole workspace under current working conditions (with product) provide the requested services. The calibration described above is carried out under real conditions on site. Optionally, the measurement can be can be carried out in a loaded or unloaded state. The measurement evaluation, including graphical representation, is made in written form. The values must not exceed the limits specified in DIN 13277:2022-05. exceeded. (Other measuring methods possible on request)

Homepage
E-mailwww.tritec-klima.com
info@tritec-klima.dePhone++49-511/3523508Fax++49-511/3521715technischeÄnderungen vorbehalten

